

# H8 Hemoglobin Analyzer (HPLC)

*<i><i>Lifotronic* 

H8

en.lifotronic.com

## H8 Hemoglobin Analyzer (HPLC)



The Fully automated Lifotronic H8 Hemoglobin Analyzer offers the fast throughput of HbA1c results in 1.2 miuntes, with Hb variant detection, providing the outstanding solution for quick and reliable diabetic monitoring. No sample preparation and very little hands-on time by the operator is required for the H8 Analyzer.

Ce	rtificate
Traceobility of Manufacturers to the IF	CC Reference Measurement Procedure for Hb.
This certifics that Shenchen Lifetronic Tes	chnology Co. Ltd. wing 119 Hh41c analyzer
	estimic thecability. In the Monhoring Programme of 201
following performance was some	
Deviation from IFCC-target	at 30 mmol HbA3cinol Hb : -1.6
	at 90 mmol HbA3c mol 11b : -4.9
Reproducibility, coefficient of	at 90 mmol IBAlicinol II5 4.9 Evariation 2.61%
Reproducibility, coefficient o Linearity, correlation coeffici	at 90 mmol IBAlicinol II5 4.9 Evariation 2.61%
	at 90 mmol IBAlicinol II5 4.9 Evariation 2.61%
Linearity, correlation coeffici	at 90 mmol HDA3ctmil Hb : -4.9 Cranistice. 3.61% ent 0.9965
Linearity, correlation coeffici	at 90-annul IDA3ctual ID : -4.9 Cranistion 3.61% and 0.9965

NGSP	Certificate of T	raceability
	Manufacturer Certific	ation
	Lifotronic Technology Co.	
certification for manufacturers and	has participated in and successfully is traceable to the Diabetes Control nation was performed with: Refere	and Complications
certification for manufacturers and i Trial Reference method. The comp for Clinical Chemistry Standards	is traceable to the Diabetes Control parison was performed with: Referen	and Complications
certification for manufacturers and i Trial Reference method. The comp	is traceable to the Diabetes Control parison was performed with: Referen	and Complications
certification for manufacturers and i Trial Reference method. The comp for Clinical Chemistry Standards The system evaluated was: Internet: IPI Benagibilis Analyzer (BPLC) Pearming	is traceable to the Dialeters Control parison was performed with: <u>Referen</u> (ReCCS) ANRL01 Colliners Let 201910281 Colliners Let 201910281 Colliners Let 201910281 Colliners Let 201910281	Column Lot P 2015 10000
certification for manufacturers and Trial Reference method. The comp for Clinical Chemistry Mandards The system evaluated was: for manufacturer (BPLK)	is traceable in the Disketes Control parton was performed with: Referen (ReCCS) ASRL#1 Control on Section Section Section (States Section Sect	Column Lo P

#### **Gold Standard of Diabetes Diagnose**

Glycosylated hemoglobin (HbA1c) is widely recognized as a Gold Standard to monitor diabetes, which can indicate the average plasma glucose concentration over 8~12 weeks.

#### **HPLC Methodology**

High-PressureLiquid Chromatography (HPLC), to separate HbA1c directly with measuring the absor bance points continually to form chromatogram. Using normal distribution curve fitting auto-iterative algorithm to get precise HbA1c testing result, excluding interference of variant and unstable hemoglobin like HbF. Standard Analysis Mode will report HbA1a、HbA1b、HbF、HbA1c、P3、HbA0 peak areas and ratio. And the result also includes IFCC, NGSP and ADAG value for diverse client needs.





Correlation between H8 and a famous HPLC method





#### **The Elements of Hemoglobin**



Product of the combination between HbA and fructose / lactose

### HPLC Technology – Gold Standard Methodology

- NGSP and IFCC certified
- HbA1c results in 1.5minutes

#### **Fully Automated - To Minimize Operation Hassles**

- Primary Tube Sampling with Cap Piercing
- Fully Automated Start-up, Maintenanceand Shutdown
- Barcode Scanner for Sample identification

#### Precise and Reliable – To Serve You Consistently

- HbA1c Inter & Intra Measuring CV's ≤1.5% to EnableExceptional ResultManagement
- SuperiorQuality Chromatographic Resolution to EliminateInterferences

#### **Compact Size - To Minimize Space Requirements**

Small Footprint Reduces Bench Space Needed



en.lifotronic.com

# H8 Hemoglobin Analyzer (HPLC)

### TECHNICAL SPECS

Methodology	High-Performance Liquid Chromatography (HPLC)
Test Modes	Fast Mode, Variant Mode, β-thalassemia Mode
Test Range	3% - 18%
Precision	CV ≤1.5%
First Sample Result	7.5 Mins
Test Speed	1.5Mins/sample for Fast Mode, 2.2 Mins/sample for Variant Mode
Sample Type	Venous Blood, Finger Peripheral Blood, Lyophilized Whole Blood
Sample Volume	6ul, Peripheral Blood, 500ul(100 Dilution Ratio)
Auto Sample Station	10 Positions
Photometer	415nm+500nm LED, 20000 Hours Life Span
Chromatography Column	Available Tests ≥1500T
Filter	≥400T
Display	10.1"TFT True Color LCD Touch Screen
Software	Embedded System with Self-Diagnosis to Monitor and Detect System Errors
Reagent Kit	Eluent A, Eluent B, Eluent C, Hemolysin, Calibrator, QC Material (Weight Sensor±1%)
Information Input	Scanner or Touch Keypad
Storage	20000 Sample Results
Connection	USB, LAN, LIS Compatible
Printer	Thermal Printer and External Laser Printer
Operation	Temperature 10~30°C (~°F)
Humidity	≤85%
Power	AC 100-240V 50/60HZ 120VA
Dimensions	580mm×500mm×520mm (22.8"H×19.7"W×20.5"D)
Weight	50kg (110lbs)
Barcode Scanner	QC Curve



#### Lifotronic Technology Co., Ltd.

Address: 4th Floor, Building 15, 1008 Songbai Road, Nanshan District, Shenzhen, 518055, China Tel: 86-755-29060197 Fax: 86-755-29060036 Email: Inter-sales@lifotronic.com Web: en.lifotronic.com